HAMI
Medical Equipment

BECKMAN COULTER

BIOCHEMISTRY

AU-5801

- Throughput: 2000 photometric tests/hour
- Reagent Position : 54 positions for R1, 54 positions for R2
- Sample Position: 400Samples
- Sample Volume : 1–17 µl in 0.1 µl steps
-13 different wavelengths between 340–800 nm

www.hamimedical.com